



magneto-electric transmitter

ES
ES-PPA
ES-FF

Ex-Supplementary Operating Manual

supplementary to the operating manuals

BGN, BGF, TSK, BA, DWF



This operating manual contains important information for the operation in potentially explosive atmospheres

Please read the instructions carefully and store them in a safe place for future reference

Table of Contents

1.	INTRODUCTION	3
I.	Shipping and storage; product inspection	3
II.	Warranty	3
III.	Application domain of the operating manual	3
IV.	Measures to be taken before sending your device to the manufacturer for repair	3
2.	STEPS PRIOR TO OPERATION	4
2.1	Installation, mounting, commissioning and maintenance	4
2.2	Hazard warnings	5
2.3	Proper use of the device	5
3.	IDENTIFICATION	6
4.	GENERAL INFORMATION ABOUT EXPLOSION PROTECTION	7
5.	SCOPE OF APPLICATION	8
6.	MODE OF OPERATION AND SYSTEM LAYOUT	8
6.1	Measuring principle	8
7.	ELECTRICAL CONNECTION	8
7.1	ES type	8
7.2	ES-PPA and ES-FF type	8
8.	EX-MARKING IN ACCORDANCE TO ATEX DIRECTIVE 2014/34/EU AND IECEx	8
9.	ELECTRICAL AND THERMAL CHARACTERISTICS	9
9.1	For the version ES	9
9.1.1	Supply and signal circuit (terminals 1 and 2)	9
9.1.2	Binary outputs 1 and 2: potential-free optocoupler output circuit (terminals 3-4 and 5-6)	9
9.2	For the versions ES-PPA (terminals 7 and 8) or ES-FF (terminals 9 and 10)	9
10.	SPECIAL CONDITIONS FOR SAFE APPLICATION	10
10.1	Environmental impacts on the transmitter	10
10.2	Atmospheric conditions	10
10.3	Ground connection	10
11.	MARKING	10
12.	DECLARATION OF CONFORMITY	11

1. Introduction

I. Shipping and storage; product inspection

The device is to be safeguarded against dampness, dirt, impact and damage.

Product inspection

Upon receipt of the product, check the contents of the box and the product particulars against the information on the delivery slip and order form so as to ensure that all ordered components have been supplied.

Notify us of any shipping damage immediately upon receipt of the product. Any damage claim received at a later time will not be honored.

II. Warranty

Your meter was manufactured in accordance with the highest quality standards and was thoroughly tested prior to shipment. However, in the event that any problem arises with your device, we will be happy to resolve the problem for you as quickly as possible under the terms of the warranty which can be found in the terms and conditions of delivery. Your warranty can only be honored if the device was installed and operated in accordance with the instructions for your device. Any mounting, commissioning and/or maintenance work is to be carried out by qualified and authorized technicians only.

III. Application domain of the operating manual



These supplementary instructions apply to explosion-proof transmitter Types ES, ES-PPA and ES-FF built **from 1st August 2016** onwards.

This manual supplements the operating manual for non-explosion proof level meters. If you do not have a copy of this manual, please request one from Heinrichs Messtechnik GmbH or download the instructions from our website. The instructions herein pertain primarily to explosion-proof level meters. The technical data in the installation and operating instructions for non-explosion proof level meters still apply insofar as the present instructions do not replace them or exclude their application.


IV. Measures to be taken before sending your device to the manufacturer for repair

It is important that you do the following before shipping your meter to Heinrichs Messtechnik GmbH for repair:

- Enclose a description of the problem with your device. Describe in as much detail as possible the application and the physical and chemical properties of the fluid.
- Remove any residues from the device and be sure to clean the seal grooves and recesses thoroughly. This is particularly important if the fluid is corrosive, toxic, carcinogenic, radioactive or otherwise hazardous.

The operator is liable for any costs arising from substance removal or personal damage due to inadequate cleaning of a device that is sent for repair.

2. Steps prior to operation



Prior to installation and operation, it is essential that the operator familiarizes himself with all of the instructions and information contained in the **manual for non-explosion proof level meters as well as the present instructions**.

If any part of either manual is missing, contact Heinrichs Messtechnik to request a new manual. These manuals can also be downloaded from our website.

2.1 Installation, mounting, commissioning and maintenance

Installation, mounting, commissioning and maintenance are to be performed by a technician trained to work with explosion-proof devices, or by a Heinrichs Messtechnik service technician.



Warning

Any maintenance or repair that safety relevant in terms of explosion-protection is to be carried out by the manufacturer, an authorized Heinrichs Messtechnik GmbH service center or under the supervision of an expert in explosion-proof devices.

Heinrichs Messtechnik GmbH accepts no liability for any loss or damage of any kind arising from improper operation of any product, improper handling or use of any replacement part, or from external electrical or mechanical effects, overvoltage or lightning. Any such improper operation, use or handling shall automatically invalidate the warranty for the product concerned.

In the event a problem arises with your device, please contact us at one of the following numbers to arrange to have your device repaired:



Phone: +49 221 49708-0



Fax: +49 221 49708-178

Internet: www.heinrichs.eu

Email: info@heinrichs.eu

Contact our customer service department if your device needs repair or if you need assistance in diagnosing a problem with your device.

2.2 Hazard warnings

The purpose of the hazard warnings listed below is to ensure that device operators and maintenance personnel are not injured and that the level meter and any devices connected to it are not damaged.

The safety advisories and hazard warnings in the present document that aim to avoid placing operators and maintenance personnel at risk and to avoid material damage are prioritized using the terms listed below, which are defined as follows in regard to the instructions herein and the advisories pertaining to the device itself:

Warning



means that failure to take the prescribed precautions **could result** in injury, substantial material damage or even death. Always comply to these warnings and proceed with caution.

Caution



means that failure to take the prescribed precaution **could result** in material damage or destruction of the device. We advise always to abide to these instructions!

Note



means that the accompanying text contains important information about the product, handling the product or about a section of the documentation that is of particular importance.

2.3 Proper use of the device



Warning

The operator is responsible for ensuring that the material used in the sensor and transmitter housing is suitable and that such material meets the requirements for the fluid being used and the ambient site conditions. The manufacturer accepts no responsibility in regard to such material and housing.



Warning

In order for the device to perform correctly and safely, it must be shipped, stored, set up, mounted operated and maintained properly.

3. Identification

Manufacturer: Heinrichs Messtechnik GmbH
Robert-Perthel-Strasse 9
D-50739 Cologne
Germany



Phone: +49 221 49708-0
Fax: +49 221 49708-178



Internet: www.heinrichs.eu
Email: info@heinrichs.eu

Product type: Magneto-electric transmitter for the angle determination of the magnetic transmission of mechanical measuring instruments.

Product name: Transmitter type: ES, ES-PPA, ES-FF

File name: es-ex_ba_20.01_en.doc

Version:. 20.01,
Date, November 11th, 2020

4. General information about explosion protection

Acc. Directive 94/9/EC (ATEX)	Example designation		CE 0158 Ex	II	2G	Ex	ia	IIC	T6	Gb
	Equipment groups									
	I	Equipment group I applies to equipment intended for use in underground parts of mines as well as those parts of surface installations of such mines endangered by firedamp and/or combustible dust.								
	II	Equipment group II applies to equipment intended for use in other places liable to be endangered by explosive atmospheres. This group is subdivided into three categories.								
Acc. EN 60079-0 ff / IEC 60079-0 ff	Equipment category									
	Gas	Dust	Definition							
	1G (0)	1 D (20)	Equipment in this category is intended for use in areas in which explosive atmospheres caused by mixtures of air and gases, vapours or mists or by air/dust mixtures are present continuously, for long periods or frequently.							
	2 G (1)	2 D (21)	Equipment in this category is intended for use in areas in which explosive atmospheres caused by gases, vapours, mists or air/dust mixtures are likely to occur.							
	3G (2)	3D (22)	Equipment in this category is intended for use in areas in which explosive atmospheres caused by gases, vapours, mists, or air/dust mixtures are unlikely to occur or, if they do occur, are likely to do so only infrequently and for a short period only.							
	(The numbers in round brackets correspond to the IEC Zones.)									
	Ex = Explosion-proof electrical equipment									
	Types of protection									
		General requirements	IEC 60079-0							
	„d“	Flameproof enclosure	IEC 60079-1							
	„e“	Increased safety	IEC 60079-7							
	„i“	Intrinsic safety (ia, ib)	IEC 60079-11							
	„t“	Equipment dust ignition protection by enclosure "t" (ta, tb or tc)	EN 60079-31							
	Explosion groups									
	Gases and vapours									
	IIA	Acetone, benzene, fuel oil, ethanoic acid								
	IIB	City gas, ethylene, isoprene								
	IIC	Acetylene, hydrogen, carbon bisulphide								
	Dust Atmospheres									
	IIIA	Fibers and flyings								
	IIIB	Non-conductive dusts								
	IIIC	Metal dusts								
	Temperature classes									
	Maximum surface temperature		Temperature class							
	450 °C	842 °F	T1							
	300 °C	572 °F	T2							
	200 °C	392 °F	T3							
	135 °C	275 °F	T4							
	100 °C	212 °F	T5							
	85 °C	185 °F	T6							
	Equipment protection level, EPL									
	Gases: Ga, Gb oder Gc		Dust: Da, Db or Dc							

Explosion protection designations [square brackets] refer to "Related electrical equipment or circuits."

5. Scope of application

The ES transducer is used in flow meters of the BGN, BGF and TSK type series in the BA fill level indicator and the DWF density meter. This covers the areas of volume flow rate measurement, fill level measurement according to the principle of positive displacement and density measurement. The ES* transducer is intended for installation in a housing providing a minimum IP class of protection from IP20.

6. Mode of operation and system layout

The electric ES-type transmitter serves to transform the needle position of the mechanical measuring system into a proportional 4-20mA signal or to the field bus interface connection.

6.1 Measuring principle

The position of the float and/or lifting body is transferred to the needle axis through a magnetic system. The ES transducer uses 2 magnetic field sensors to measure the field of a magnet attached to the needle axis in order to generate an output current of 4...20 mA. The usual non-linear scale is in this case linearised with a maximum of 16 data points.

The terrestrial magnetic field and moderately sized homogenous external magnetic fields are largely compensated by the applied differential measurement of the 2 magnetic field sensors.

7. Electrical connection

7.1 ES type

The electrical connection of the ES type is provided through an intrinsically safe 2-wire supply and signal circuit of 4-20 mA.

7.2 ES-PPA and ES-FF type

The ES-PPA and/or ES-FF types are "FISCO field devices" and the electrical connection is realised via an intrinsically safe 2-wire field bus circuit according to the FISCO model.

As an option, devices may also be connected to intrinsically safe field bus circuits that do not correspond to the FISCO model. In this case careful attention must be paid to the maximum electrical values (U_i , I_i , P_i , L_i and C_i) as described below.

8. Ex-Marking in accordance to ATEX directive 2014/34/EU and IECEx



DMT 00 ATEX E 075
II 2G Ex ia IIC T6 Gb



IECEx BVS 16.0072
Ex ia IIC T6 Gb

9. Electrical and thermal characteristics

9.1 For the version ES

9.1.1 Supply and signal circuit (terminals 1 and 2)

Voltage	U_i	DC	30	V
Current	I_i		150	mA
Power	P_i		1	W
Effective internal inductance	L_i		0,24	mH
Effective internal capacitance	C_i		16	nF

9.1.2 Binary outputs 1 and 2: potential-free optocoupler output circuit (terminals 3-4 and 5-6)

each

Voltage	U_i	DC	30	V
Current	I_i		20	mA
Power	P_i		100	mW
Effective internal inductance	L_i		4	μ H
Effective internal capacitance	C_i		16	nF
Ambient temperature range	T_a		-40 °C to + 70 °C	

9.2 For the versions ES-PPA (terminals 7 and 8) or ES-FF (terminals 9 and 10)

For the use as a field device in an intrinsically safe field bus system in acc. FISCO (IEC 60079-11, Annex G), or for the connection to intrinsically safe electric circuits.

Parameters for the transmitter:

$$U_i = 32 \text{ V}$$

$$I_i = 280 \text{ mA}$$

$$P_i = 2 \text{ W}$$

$$C_i < 5 \text{ nF}$$

$$L_i < 10 \mu\text{H}$$

The ambient temperature range amounts to -40 °C to +70 °C.

10. Special conditions for safe application

10.1 Environmental impacts on the transmitter.

Environmental impacts, like the process temperature of the flow meter and the installed ES* transmitter must be considered. See also item 7 of the measuring device's general instruction manual.

10.2 Atmospheric conditions

In accordance with EN 1127, a "potentially explosive atmosphere" is defined as a mixture of air and combustible gases, vapour, mist or dust under atmospheric conditions. Such conditions are defined in EN 13463-1, para. 1, with values $T_{atm} = -20\text{ °C}$ to $+60\text{ °C}$ and $P_{atm} = 0.8$ to 1.1 bar. Outside this range, safety parameters for most ignition sources are not available.

Usually, variable-area flow meters operate under operating conditions outside the atmospheric conditions of 0.8 to 1.1 bar. Irrespective of the zone classification –safety parameters of explosion protection – are basically not applicable to the inside of the measuring tube.

Therefore operation with combustible products is only allowed if a potentially explosive air mixture is not formed inside the flow meter. Where this condition is not met, the operator will need to assess the ignition hazard in each individual case and give due consideration to existing parameters (e.g. pressure, temperature, process product, materials of construction for the measuring tube).

10.3 Ground connection







In variable-area flow meters, it is possible under operating conditions for charge separation to occur in the measuring tube due to the transport of non-conductive fluids and/or when the flow comes into contact with non-conductive internals (e.g. liners, floats). For that reason, variable-area flow meters must be permanently grounded by the operator by way of the process connections (flanges) in order to discharge electrostatic build-up. The operator is also responsible for extending the ground continuity of the process pipeline. If grounding cannot be made via the process connections (plastic process connections or undefined connections), the flow meter must be connected to the local ground potential via the flanges. This connection only ensures electrostatic grounding of the device and does not meet the requirements for equipotential bonding.

11. Marking



Only devices bearing the Ex label may be operated in explosive environments.

Example type plates for the ES standard and ES-FF versions are presented below.

 D-50739 Köln Robert-Perthel-Str. 9 www.heinrichs.eu	 0158  Type: ES Tamb: -40°C to +70°C MF-Date: 2017 Ser. No.: 123456	DMT 00 ATEX E 075 IECEx BVS 16.xxxx II 2G Ex ia IIC T6 Gb Terminals 1 and 2 U _i = 30 VDC, I _i = 150 mA P _i = 1 W L _i = 0,24 mH, C _i = 16 nF	 D-50739 Köln Robert-Perthel-Str. 9 www.heinrichs.eu	 0158  Type: ES-FF Tamb: -40°C to +70°C MF-Date: 2017 Ser. No.: 123456	DMT 00 ATEX E 075 IECEx BVS 16.xxxx II 2G Ex ia IIC T6 Gb FISCO Field Device Terminals 9 and 10 U _i = 32 VDC, I _i = 280 mA P _i = 2 W L _i = 10 µH, C _i = 5 nF
---	--	---	---	---	--

12. Declaration of Conformity

EU-Konformitätserklärung **EU-Declaration of Conformity**



Nº. 20-4148-01

Hersteller:
Manufacturer: Heinrichs Messtechnik GmbH
 Robert-Perthel-Strasse 9
 50739 Köln

Produktbeschreibung:
Product description: **Messwertumformer ES bzw. ES-PPA und ES-FF zur Erfassung der Position oder Winkelstellung eines Magnets bei Schwebekörpern-Durchflußmessgeräte.**
Electronic transmitter ES, ES-PPA or ES-FF for the position acquisition of magnets in variable area flowmeters.

Hiermit erklären wir, in alleinige Verantwortung, dass das oben genannte Messsystem den Anforderungen der folgenden EU-Richtlinien, einschließlich allen bis heute veröffentlichten Änderungen bzw. Nachträgen entspricht:

We declare herewith, in sole responsibility, that the product described above is conform with the provisions of the following EU-directives, including all published changes and amendments as of today:

- | | |
|--------------------------|--|
| 2014/30/EU (EMC) | EU-Richtlinie über die Elektromagnetische Verträglichkeit
<i>EU-Directive relating to electromagnetic compatibility</i> |
| 2014/34/EU (ATEX) | EU-Richtlinie über Geräte zur Bestimmungsgemäße Verwendung in explosionsgefährdeten Bereichen.
<i>EU-Directive relating to electrical equipment intended for use in potentially explosive atmospheres</i> |
| 2014/35/EU (LVD) | EU-Richtlinie über die Bereitstellung elektrischer Betriebsmittel zur Verwendung innerhalb bestimmter Spannungsgrenzen auf dem Markt
<i>EU-Directive on the harmonisation of the laws of the Member States relating to the making available on the market of pressure equipment</i> |

Anhang N und X sind ein integraler Bestandteil dieser Erklärung
Annex N and X are an integral part of this declaration

Köln, den 01.04.2020



Joseph Burke
 (Explosionsschutzbeauftragter /
 Explosion Protection Representative)



Guido Thometzki
 (Geschäftsführung / Managing Director)

Kontakt: Tel: +49 (221) 49708-0
Contact: Email: info@heinrichs.eu
 Web: www.heinrichs.eu

Konformitätserklärung-ES_20-4148-01.docx

Seite 1 von 3



Anhang N zur EU-Konformitätserklärung Annex N of the EU-Declaration of Conformity



Nº. 20-4148-01

Produktbeschreibung:
Product description:

Messwertumformer ES bzw. ES-PPA und ES-FF zur Erfassung der Position oder Winkelstellung eines Magnets bei Schwebekörpern-Durchflußmessgeräte.
Electronic transmitter ES, ES-PPA or ES-FF for the position acquisition of magnets in variable area flowmeters.

Durch diese Erklärung wird die Konformität mit den auf Seite 1 genannten Richtlinien sowie die Einhaltung der folgenden Normen bestätigt (gegebenenfalls abhängig von Gerätevariante):
Conformity to the Directives referred to on Page 1 of this Declaration is assured through the application of the following standards (possibly dependent on version of device):

Richtlinie Directive	Norm –Ref. Nr. Standard / Ref. Nº.	Ausgabe Edition	Norm Beschreibung Standard Description	ES	ES-PPA	ES-FF
	DIN EN -					
2014/30/EU	61000-6-2	2011-06	Immunity Industry environment	X	X	X
	61000-6-3	2012-11	Emission residential environment	X	X	X
	55011	2011-04	Radio frequency disturbance	X	X	X
	61326-1	2011-07	EMC requirements	X	X	X
2014/34/EU	60079-0	2012+ A11:2013	General requirements	X	X	X
	60079-11	2012	Intrinsic Safety „i“	X	X	X
2014/35/EU	61010-1	2011	Safety requirements	X	X	X

X: Zutreffende Norm / Applicable Standard

Name und Anschrift der Notifizierte Stelle / Name and Address of the Notified Body

TÜV-SÜD Industrie Service GmbH
TÜV SÜD Gruppe
Westendstraße 199
D-80686 München

DEKRA Testing and Certification GmbH
Carl-Beyling-Haus
Dinnendahlstraße 9
D-44809 Bochum
ID-Nr. / ID-Nº.: RL 2014/34/EU: 0158





Anhang X zur EU-Konformitätserklärung Annex X to the EU-Declaration of Conformity



Nr. 20-4148-01

Produktbeschreibung: **Messwertumformer ES bzw. ES-PPA und ES-FF zur Erfassung der Position oder Winkelstellung eines Magnets bei Schwebekörpern-Durchflußmessgeräten.**
 Product description: **Electronic transmitter ES, ES-PPA or ES-FF for the position acquisition of magnets in variable area flowmeters.**

Gerät Zulassungen / Device certification

Prüfbescheinigungen examination certificates	Nachtrag Supplement	Kennzeichnung Marking	ES	ES-PPA	ES-FF
DMT 00 ATEX E 075	2	II 2G	X	X	X

X: Zutreffende Norm / Applicable Standard

Die oben genannten Produkte entsprechen der Richtlinie 2014/34/EU. Neue Editionen können bereits eine oder mehrere der in den jeweiligen EU-Baumusterprüfbescheinigungen genannten Normen ersetzt haben. Der Hersteller erklärt, dass alle in dieser Konformitätserklärung erwähnten Produkte auch die Anforderungen der neuen Ausgaben einhalten, da die veränderten Anforderungen der neuen Ausgaben entweder keinen Einfluss auf das Produkt haben, oder das Produkt die Anforderungen erfüllt.

The above-mentioned products comply with the Directive 2014/34/EU. New editions may have already replaced one or more of the Standards stated in the respective EU-Type-examination certificates. The manufacturer declares that all products mentioned in this Declaration of Conformity also comply with the requirements of the new editions since either the changed requirements of the new editions do not affect the product, or the product also fulfills the requirements.

Heinrichs Messtechnik GmbH

Robert-Perthel-Straße 9
50739 Köln
Telefon 0221/49708-0
Telefax 0221/49708-178
<http://www.heinrichs.eu>
info@heinrichs.eu

Bankverbindung

Dresdner Bank Köln
BLZ 370 800 40
Konto-Nr. 0955 051300
IBAN :
DE58 3708 0040 0955 0513 00
SWIFT-BIC: DRES DE FF 370

Erfüllungsort und Gerichtsstand:

Köln
Amtsgericht Köln HRA 37040

Ust.IDNr.: DE813416533
Steuer-Nr.: 217/5743/0386

Geschäftsführer

Dipl. Ing. (FH)
Guido Thometzki

Konformitätserklärung-ES_20-4148-01.docx

Seite 3 von 3