



# EXPLOSION PROTECTION CERTIFICATE OF CONFORMITY

Cert NO.GYJ17.1173

This is to certify that the product

Electronic transmitter

manufactured by Heinrichs Messtechnik GmbH

(Address:Robert-Perthel-Strasse 9, 50739 Cologne, Germany)

which model is ES, ES-PPA, ES-FF

Ex marking Ex iaIIC T6 Gb

product standard /

drawing number ES-Ex-01 a

has been inspected and certified by NEPSI, and that it conforms to GB 3836.1-2010,GB 3836.4-2010

This Approval shall remain in force until 2022.06.14

Remarks 1.Conditions for safe use are specified in the attachment(s) to this certificate.  
2.Intrinsic safety parameters specified in the attachment(s) to this certificate.

Director

National Supervision and Inspection Centre for  
Explosion Protection and Safety of Instrumentation

Issued Date 2017.06.15

This Certificate is valid for products compatible with the documents and samples approved by NEPSI.

103 Cao Bao Road  
Shanghai 200233, China

<http://www.nepsi.org.cn>  
Email: [info@nepsi.org.cn](mailto:info@nepsi.org.cn)

Tel: +86 21 64368180  
Fax: +86 21 64844580



# 防爆合格证

证号: GYJ17.1173

由 Heinrichs Messtechnik GmbH

制造的产品:

(地址: Robert-Perthel-Strasse 9, 50739 Cologne, Germany)

名称 磁电变送器

型号规格 ES, ES-PPA, ES-FF

防爆标志 Ex ia II C T6 Gb

产品标准 /

图样编号 ES-Ex-01 a

经图样及技术文件的审查和样品检验, 确认上述产品符合 GB 3836.1-2010、GB 3836.4-2010 标准, 特颁发此证。

本证书有效期: 2017年6月15日至2022年6月14日

备注 1. 安全使用注意事项见本证书附件。  
2. 本安电气参数见本证书附件。

站长

国家级仪器仪表防爆安全监督检验站

颁发日期二〇一七年六月十五日



本证书仅对与认可文件和样品一致的产品有效。

地址: 上海市漕宝路103号  
邮编: 200233

网址: [www.nepsi.org.cn](http://www.nepsi.org.cn)  
Email: [info@nepsi.org.cn](mailto:info@nepsi.org.cn)

电话: +86 21 64368180  
传真: +86 21 64844580