



EXPLOSION PROTECTION CERTIFICATE OF CONFORMITY

Cert NO:GYJ17.1167X

This is to certify that the product

Coriolis Flowmeter Converter (with control unit BE2)

manufactured by Heinrichs Messtechnik GmbH

(Address:Robert-Perthel-Strasse 9, 50739 Cologne, Germany)

which model is UMC3 Series

Ex marking Ex d [ia Ga]IIB/IIC T6 Gb, Ex d e [ia Ga]IIB/IIC T6 Gb

product standard /

drawing number HBR-EX-002 c

has been inspected and certified by NEPSI, and that it conforms to GB 3836.1-2010,GB 3836.2-2010,GB 3836.3-2010,GB 3836.4-2010, GB 3836.20-2010

This Approval shall remain in force until 2022.06.14

Remarks

- 1.Conditions for safe use are specified in the attachment(s) to this certificate.
- 2.Symbol "X" placed after the certification number denotes specific conditions of use, which are specified in the attachment(s) to this certificate.
- 3.Model designation is specified in the attachment(s) to this certificate.
- 4.Safe parameters specified in the attachment(s) to this certificate.

Director

National Supervision and Inspection Centre for
Explosion Protection and Safety of Instrumentation


Issued Date 2017.06.15

This Certificate is valid for products compatible with the documents and samples approved by NEPSI.

103 Cao Bao Road
Shanghai 200233, China

<http://www.nepsi.org.cn>
Email: info@nepsi.org.cn

Tel: +86 21 64368180
Fax: +86 21 64844580



防爆合格证

证号: GYJ17.1167X

由 Heinrichs Messtechnik GmbH

制造的产品:

(地址: Robert-Perthel-Strasse 9, 50739 Cologne, Germany)

名称 质量流量计转换器 (带控制器BE2)

型号规格 UMC3系列

防爆标志 Ex d [ia Ga] II B/II C T6 Gb、Ex d e [ia Ga] II B/II C T6 Gb

产品标准 /

图样编号 HBR-EX-002 c

经图样及技术文件的审查和样品检验, 确认上述产品符合 GB 3836.1-2010、GB 3836.2-2010、GB 3836.3-2010、GB 3836.4-2010、GB 3836.20-2010 标准, 特颁发此证。

本证书有效期: 2017年6月15日至2022年6月14日

- 备注
1. 安全使用注意事项见本证书附件。
 2. 证书编号后缀“X”表明产品具有安全使用特殊条件, 内容见本证书附件。
 3. 型号规格说明见本证书附件。
 4. 电气安全参数见本证书附件。

站长

国家级仪器仪表防爆安全监督检验站

颁发日期二〇一七年六月十五日

本证书仅对与认可文件和样品一致的产品有效。

地址: 上海市漕宝路103号
邮编: 200233

网址: www.nepsi.org.cn
Email: info@nepsi.org.cn

电话: +86 21 64368180
传真: +86 21 64844580